

CERTIFICATE OF ANALYSIS

Page 1 of 1

CUSTOMER

Addmaster Ltd
Darfin House
Priestly Court
Staffordshire Technology Park
Stafford
ST18 0AR
UK.

CERTIFICATE NO.

1038185.40/13224

CUSTOMER REF.

9100

SAMPLE DETAILS

CBS

DATE RECEIVED

23/07/2019

ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

DATE ANALYSED

24/07/2019

DATE REPORTED

26/07/2019

RESULTS (AS CFU CM⁻²)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (CONTROL)	
		0 hrs	24 hrs	Log 10	%
MEDICAL MAT (CONTROL)	<i>E coli</i>	1.7E+04	1.2E+05		
MEDICAL MAT WITH BIOMASTER (AT300 @ 0.3%)	<i>E coli</i>	1.7E+04	< 11.11	≥ 4.04	≥ 99.99%
MEDICAL MAT (CONTROL)	MRSA	1.9E+04	7.1E+03		
MEDICAL MAT WITH BIOMASTER (AT300 @ 0.3%)	MRSA	1.9E+04	< 11.11	≥ 2.81	≥ 99.84%

Key: NS = Poor survival on control supplied.

The above data show the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the control sample.

IMSL MICROBIOLOGICAL SERVICES LTD
PALE LANE
HARTLEY WINTNEY
HANTS RG27 8DH
UK

MANAGING DIRECTOR
Peter D Askew

